AMENDMENT NO. 3 AUGUST 2022

TO

IS 16112 : 2013 BETA PICOLINE — SPECIFICATION

[*Page* 2, *Annex* A, *clause* **A-2.3** *(see also Amendment No*. 2)] — Substitute the following for the clause **A-2.3**.

'**A-2.3 Column —** General purpose nonpolar capillary GC column containing 100 percent dimethyl polysiloxane (PDMS) or capillary GC column containing high polarity polyethylene glycol (PEG) phase, 60 m long, 0.25 mm internal diameter, 0.25 µm film thickness (df).

NOTE — The above gas chromatographic conditions are suggestive. However any GC having different columns (packed/Capillary having different length/diameter/film thickness) and different carrier gas (He, H_2 or N_2), with different calibration technique (internal standard, external standard, area normalization) may be used, provided standardization/calibrations are done after setting up chromatographic conditions for required resolution.'

(PCD 09)

Publication Unit, BIS, New Delhi, India